



eurofins



TEST CERTIFICATE

CML-TC20210616.18

TEST FOR THE INGRESS PROTECTION OF A Type 11 Connection Head

Manufacturer:	H&B Sensors Ltd Odyssey House Durban Road Bognor Regis West Sussex PO22 9RH United Kingdom
Model Identification:	Type 11 Connection Head
CML Sample Reference:	S14072-1
Objective:	IP66 and IP68 (2 metres for 2 hours)
Standard:	EN 60529:1991 +A2:2013 IEC 60529:1989 +A2:2013
CML Test Report:	CML-TR20210616.18
Sample Delivered:	26 th May 2021
Test Dates:	18 th June to 28 th June 2021

This certificate confirms compliance of the sample tested against the agreed test plan.

Compliance of other samples cannot necessarily be inferred.

Certification Management Limited in accordance with ISO/IEC 17025 has tested the sample listed and has found that it complies with Ingress Protection Code: IP66 and IP68 (2 metres for 2 hours).

M.Powers
Senior Laboratory Engineer

Dated: 8th July 2021

This test certificate shall not be reproduced, except in full, without the written approval of CML